

*pH Sensor*  
**PH1000 User Manual**

*Version 1.1*



*PH1000 User Manual*

# Table of Contents

---

<b>1</b>	<b>Introduction .....</b>	<b>4</b>
1.1	System Description .....	4
1.2	How to Use the Manual .....	4
1.3	Certification .....	5
1.4	Unpacking and Inspection .....	5
1.5	Serial Number .....	5
1.6	Warranty Policy .....	6
1.7	Factory Service & Repair .....	6
1.8	Contact Details .....	7
<b>2</b>	<b>Sensor Overview .....</b>	<b>8</b>
2.1	Theory of Measurement .....	8
2.2	Applications .....	8
2.3	Instrument Details .....	9
	Sensor Design .....	9
2.4	Cable Details .....	10
	Mechanical Specifications .....	10
2.5	Options and Accessories .....	11
	Sensor Body .....	11
	Cable Options (Inc Detachable Cables) .....	11
	Replacement Electrode Kit (Part #570-0230) .....	11
	Accessory Kit (Part # 5PH-AK1) .....	11
2.6	Sensor Factory Calibration .....	11
<b>3</b>	<b>Sensor Operation .....</b>	<b>12</b>
3.1	Wiring & Connections .....	12
3.2	Calibration Quick Check .....	13
	Quick check .....	13
	Calculating Gain and Offset Correction Factors .....	13
3.3	Installation .....	14
	Shipping and Storage Cap .....	14
	Field Deployment Considerations .....	14
	Cabling Considerations .....	15
	Typical Sensor Installations .....	15
	Field Installation must ensure: .....	15
	Other Considerations .....	15
	Guidelines for cleaning equipment .....	16
	Replacing Reference Protection Ring .....	16
	Electrode storage .....	16
	Electrode replacement .....	17

<b>4</b>	<b>Appendix A -Additional Information .....</b>	<b>18</b>
4.1	Specifications.....	18
4.2	Quick Start Guide – PH1000 Sensor.....	19
4.3	Certificate of Conformance.....	20
4.4	Engineering Note – Detachable Cables .....	21

# 1 Introduction

## 1.1 System Description

Thank you for purchasing the Greenspan pH Sensor Model PH1000. This manual provides a guide to the configuration, operation and maintenance of the sensor to provide long term reliable and accurate monitoring.

The PH1000 probe uses a combination, glass pH and reference electrode in the one body. This electrode is field replaceable by the user. When placed in an aqueous solution a voltage is developed in the pH electrode which varies depending on the pH of the solution. This voltage is compared against the voltage from the reference electrode which is stable. The reference consists of a specially coated wire in a salt filled cavity. This cavity makes electrical contact with the outside aqueous solution through a small porous wick. A common cause for premature pH sensor failure is the chemical or biological contamination or fouling of the wick or salt filled cavity. Added protection for the reference electrode is provided by a reference protection ring constructed of porous PTFE that is impregnated with a special conductive gel. It is placed over the reference wick and should prevent most contaminants from reaching the wick, while still allowing good electrical contact with the sample. If the low cost protection ring is contaminated or fouled it can be simply replaced, extending the life of the combination electrode.

Designed for applications where an analogue output is required. The sensor provides a 3 wire 4-20mA output. The sensor requires power from a stable DC source between 9 and 16 Volts.

The Sensor is packaged in an Acetal or 316 Stainless Steel tube, with Double O ring connections for the cable and sensor head. This fully submersible sensor design is rugged and well proven and can withstand the harsh conditions found in remote field applications including groundwater, salty or acidic water conditions.

## 1.2 How to Use the Manual

Along with this manual, there are several other documents that may assist in the successful configuration and operation of the Greenspan PH1000 Sensor. These should be maintained on file as a permanent reference as to the features, applications and use of the PH1000.

***Greenspan PH1000 – Specifications Brochure***

***Greenspan PH1000 – Certificate of Conformance***

***Greenspan PH1000 – Quick Start Guide***

## 1.3 Certification

The pH probes are assembled and tested in accordance with Greenspan's ISO 9001 Quality Certified System. Each Sensor is individually manufactured and certified against a traceable Standard ([See Section 2.6](#))

Following calibration the sensors undergo a range of additional control processes to ensure that all specifications are consistent and documented.

- ***The instrument is visually inspected, marked and labelled.***
- ***The complete sensor calibration record is archived for reference, and batch number information is kept on file for statistical analysis.***
- ***An individual Certificate of Conformance is issued to the customer.***

## 1.4 Unpacking and Inspection

All Greenspan Analytical Sensors are made to order and are individually calibrated and inspected. This ensures that they leave the factory in a working condition. They are packed in new cartons for shipping. On receipt, the customer should inspect the packaging and contents for any signs of damage during transportation. The customer should also check that all items on the delivery note have been received.

Please contact the factory in case anything has been damaged or missing. A full set of documentation including Certificate of Conformance, Quick Start Guide, and Full Operator Manual will be provided with all equipment – either in hard copy format or in electronic format on the CD shipped with the goods.

If fitted with a 316 Stainless Steel body, the unit should only be used in low EC situations. Care should be taken against possible corrosion in high Chloride or Ferric solutions, water with high iron or sulphate reducing bacteria, or low dissolved oxygen. The Sensor can be fitted with an Acetal body which provides superior corrosion protection in a wide range of chemically active waters.

Because an individual sensor may be used in a variety of locations, media compatibility should be checked before installing and advice sought from Greenspan if any doubt exists.

## 1.5 Serial Number

Checking the Model Number and Range

Before installing your Greenspan PH1000 sensor check the information on the label is correct to confirm you have received the instrument you have ordered. The label will look similar to this.

<b>MODEL</b>	<b>PH1000</b>
<b>RANGE</b>	<b>0 – 14pH</b>
<b>S/N</b>	<b>012345</b>

The customer is advised to keep a record of the serial numbers in case the sensor is lost or the label damage. Greenspan Analytical keeps records of all sensors sold including a calibration history.

## 1.6 Warranty Policy

Greenspan Analytical warrants all new Greenspan products against defects in materials and workmanship for **12 months** from the date of invoice.

Products that prove to be defective during the warranty period will be repaired or replaced at the discretion of Greenspan Analytical.

Under Greenspan Analytical warranty conditions; it is the responsibility of the customer to cover shipping charges back to the factory. Upon repair/replacement Greenspan Analytical will cover the return shipping charges to the customer.

This warranty does not apply to products or parts thereof which have been altered or repaired outside of the Greenspan Analytical factory or other authorised service centre; or products damaged by improper installation or application, or subjected to misuse, abuse neglect or accident. This warranty also excludes items such as reference electrodes and Dissolved Oxygen membranes that may degrade during normal use.

Greenspan Analytical will not be liable for any incidental or consequential damage or expense incurred by the user due to partial or incomplete operability of its products for any reason whatsoever or due to inaccurate information generated by its products.

All Warranty service will be completed as soon possible. If delays are unavoidable customers will be contacted immediately.

Any sensor should not be dismantled unless under instruction from Greenspan Analytical Technical Service staff. Incorrect handling will void the warranty.

## 1.7 Factory Service & Repair

The correct choice of sensor and assistance with field installation can be provided by Greenspan and their sales offices. A correct choice of equipment, together with technical advice and field experience should result in long term success in the field. **Greenspan Technical Services** is dedicated to customer support and provides assistance in the selection, installation, deployment and commissioning of sensors with a full range of consulting services. All Greenspan products are designed, developed and manufactured in Australia and can be supplied at short notice.

If for some reason sensors are required to be returned to our factory or your sales representative, please note the model and serial number, Describe the problem, including how and under what conditions the instrument was being used at the time of malfunction. Clean the product and cable. Decontaminate thoroughly if used in toxic or hazardous environment. Carefully pack product in original packaging if possible & include a statement certifying product and cable have been decontaminated with supporting information. Products returned for repair must be accompanied by a completed GRA (Goods Return Advice) form. All sensors returned for service and repair work must be properly decontaminated prior to return. A cleaning charge may be applied to sensors that require further decontamination. Service work will not commence until the quotation has been accepted by the customer. A purchase order for all repair and service work will be required before work is carried out.

## 1.8 Contact Details

### Australia

**Head Office**  
Goyen Controls Co Pty Ltd  
268 Milperra Road  
Milperra, NSW 2214

Telephone: 1800 805 372  
Facsimile: 1300 658 799

### Sales and Service

**Queensland, South Australia  
Victoria, Western Australia**

Telephone: 1800 805 372  
Facsimile: 1300 658 799

### USA

Goyen Valve Corporation  
1195 Airport Road  
Lakewood  
New Jersey 08701, USA

Telephone: 1 732 364 7800  
Facsimile: 1 732 364 1356

### Asia

Goyen Controls Co Pty Ltd  
Shanghai Representative Office  
1209 Greenland Business Centre  
1258 Yu Yuan Road  
Shanghai PC200050, CHINA

Telephone: 86 21 5239 8810  
Facsimile: 86 21 5239 8812

Goyen Controls Co Pty Ltd  
73-M Jalan Mega Mendung  
Kompleks Bandar OUG  
58200 Kuala Lumpur, MALAYSIA

Telephone: 60 37 987 6839  
Facsimile: 60 37 987 7839

Greenspan Singapore Pte Ltd  
02-01, Minwa Industrial Building  
39 Genting Lane  
Singapore 349554

Telephone: 65 6748 0140  
Facsimile: 65 6748 2534

### Europe

Goyen Controls Co UK Ltd  
Unit 3B Beechwood  
Chineham Business Park  
Basingstoke, Hampshire, RG24 8WA  
UNITED KINGDOM

Telephone: 44 1256 817 800  
Facsimile: 44 1256 843 164

**Tyco Umwelttechnik GmbH**  
Im Petersfeld 6  
D-65624 Altdenz  
GERMANY

Telephone: 49 6432 1001/1002  
Facsimile: 49 6432 63810

**Mecair S.r.l.**  
Via per Cinisello 97  
20054 Nova Milanese  
Milano,  
ITALY

Telephone: 39 362 375 118  
Facsimile: 39 362 375 124

**Address:** *Tyco Environmental Systems  
Greenspan Analytical Manufacturing Plant  
22 Palmerin Street  
WARWICK QLD 4370  
AUSTRALIA*

**Phone:** + 61 (0)7 46601888  
**Fax:** + 61 (0)7 46601800

**Internet:** [www.tyco-environmental.com](http://www.tyco-environmental.com)

## 2 Sensor Overview

### 2.1 Theory of Measurement

The Greenspan PH 1000 sensor uses a robust, gel filled, industrial pH electrode for field monitoring in a variety of environments.


The pH electrode consists of a PH sensitive glass membrane attached to a sealed insulating tube containing a solution of fixed PH in contact with a silver-silver chloride half cell. The potential developed across the membrane is compared to a stable reference potential e.g. a silver-silver chloride half cell in contact with an electrolyte containing chloride. Completion of the circuit is by means of a porous constriction (the salt bridge) which allows the reference electrolyte to slowly flow into the sample.

pH gives an indication of the acidity/alkalinity of a solution and is defined as:

$$pH = -\log (H^+)$$

and covers a scale from 0 (acid) to 14 (alkaline) where  $H^+$  is the hydrogen concentration in solution, at ordinary temperatures.

EG. pH of water



The concentration of each type of ion is  $10^{-7}$ gm molecule/litre and hence the pH of pure water is:

$$pH = -\log 10^{-7} = 7$$

### 2.2 Applications

Applications in which the Greenspan PH1000 can be used include:


- Monitoring of streams and rivers.
- Monitoring of water storage bodies including stratification studies.
- Intermediate and final effluent treatment monitoring.
- Hydrological run off studies.
- Ground and bore water analysis.
- Drinking water filtration efficiency.
- Industrial process monitoring.
- Sludge and dredge monitoring.

## 2.3 Instrument Details

### Sensor Design


The Greenspan pH Sensor consists of the following primary elements:

- Microprocessor controlled signal conditioning and logging device
- Stainless steel or Acetal Body Material
- Sensor Head Shroud (removable for cleaning)



## 2.4 Cable Details

All Greenspan Sensors utilise a specially designed Polyurethane Cable. The cable contains 12 x conductors, 1 x drain wire, and an internal vent tube. The outer jacket is made from UV stabilized Polyurethane and is suitable for all external, underwater or harsh environment applications. This common cable construction is utilized for vented and non vented sensors and all Greenspan Water Quality Sensors. Cables are generally factory fitted at time of manufacture in specified lengths. Cables can be joined or repaired in the field providing a waterproof connection can be maintained. Alternatively, cables can be terminated in waterproof junction boxes where cabling to other devices or longer cable runs are required.



### Cable Construction

- 1 – **Vent Tube:** Polyamide  
(size ID x OD) 2.40 x 3.20 mm
- 2 - **12 x Conductors**  
7 x 0.20 mm Tinned Copper  
Section = 0.22mm<sup>2</sup> AWG24  
Insulation: Polypropylene  
(size) = 1.10 mm ± 0.05 mm
- 3 - **Tape:** Polyester
- 4 – **Drain Wire:** 7x0.20 TinCu
- 5 - **Tape:** Polyester Aluminium
- 6 - **Jacket:** Polyurethane black,  
(size OD) 8.05 mm ± 0.15

### Mechanical Specifications

- Specially Manufactured Greenspan Cable with 12 cores and Internal Vent
- High chemical resilience and abrasive resistance
- Conductor cross section : AWG 24,
- Electrical Resistance 9 ohm per 100m (per conductor)
- Operating temperature: 85°C (max.),
- Bending radius (static) : 6 ,
- Bending radius (dynamic) 12.
- Max Operating voltage : 250V
- Jacket Printing (white colour each meter)
- Conductor colour codes : green, yellow, white, black, brown, turquoise, violet, pink, red, blue, grey
- Tensile Strength is sufficient to self suspend the Greenspan Sensor to depths of 300m.
- Long term creep due to temperature effects or tensile loading is negligible.

The moulded cable is fitted to the sensor using a double o ring seal and located using 2 x grub screws. The length of the cable is not critical to the long term calibration and operation of the sensor (provided the electrical requirements such as minimum supply voltage are maintained).

## 2.5 Options and Accessories

### **Sensor Body**

Sensor is available with a black Acetal or passivated 316 Stainless Steel body. For applications in harsh environments it is recommended that the Acetal body material be specified.

### **Cable Options (Inc Detachable Cables)**

A standard sensor is supplied with a fixed moulded cable entry. The cable is supplied with bare wires at the end opposite to the sensor.

As an option, your sensor can be ordered with a detachable cable. Detachable cables are available in a range of standard lengths and are interchangeable amongst the range of Greenspan sensors. This option can provide benefits and cost savings. Detachable cables are supplied with a 7 pin Hirschman connector at the end opposite to the sensor. An adapter cable with a mating 7 socket Hirschman connector broken out to bare wires is available (Part # 5CC-750).

Please refer to the [Engineering Note](#) in the appendix section of the manual for detailed instructions on connecting and disconnecting the detachable cable.

### **Replacement Electrode Kit (Part #570-0230)**

A kit containing a replacement combination electrode, spare reference protection ring and cleaning tool is available. Contact your Greenspan Analytical sales representative for pricing and details.

### **Accessory Kit (Part # 5PH-AK1)**

A kit containing a reference protection ring and cleaning tool is available. Contact your Greenspan Analytical sales representative for pricing and details.

## 2.6 Sensor Factory Calibration

- Sensors are calibrated using pH buffers.
- An extensive range of final calibration and inspection tests are carried out on every sensor.
- The sensor is visually inspected and packed, ready for despatch.
- The complete calibration records, sensor history and batch number are placed on file and archived.

## 3 Sensor Operation

### 3.1 Wiring & Connections

The PH1000 is a 3 wire, 4-20mA output sensor.


Red – 10 -14 VDC+

Blue – 4-20 mA output

Green – Power and signal ground.

There is also provided a green/yellow wire that connects to the cable shield.

**The following diagram illustrates the typical wiring arrangement for the PH1000**



## 3.2 Calibration Quick Check

The Factory supplied Calibration Certificate will provide detailed information on the calibration of the sensor and should be retained for future reference.

To maintain high quality control over monitoring programs, it is recommended pH calibration is checked every 3-6 months. The PH1000 has no facility to adjust the analogue outputs. If any gain or offset correction is required the factors should be entered into the datalogger or measuring device. Alternatively sensors may be returned to an authorised Greenspan agent for re-calibration.

The sensor can be checked by use of known pH standards (buffer solutions).

### Quick check

1. Remove the sensor from the water, ensure the sensor is clean. Dirty electrodes could be a key source of calibration errors.
2. Connect the sensor to a DC power source and a meter (ideally the measuring device the sensor is connected to in its application) to measure the sensor output.
3. Place the sensor in a low level pH buffer (eg. pH 4) so that the electrode is fully submerged. Record the measured output value.
4. Carefully rinse the electrode in water and shake off any attached droplets.
5. Place the sensor in a high level pH buffer (eg. pH 10) so that the electrode is fully submerged. Record the measured output value.
6. The pH sensor should output a value that is proportional to the full scale range of the sensor +/- 3% of the full scale range of the sensor (eg. +/- 0.48mA).
7. This confirms that the sensor electronics has remained stable and no further action should be required if the sensor is within +/- 3 % FS.

### Calculating Gain and Offset Correction Factors

Most dataloggers or SCADA devices have the ability to enter gain (or multiplier) and offset factors to correct for small changes in the sensor output. If the quick check values are outside allowable tolerances, correction factors can be calculated and entered into the measuring device or used to post process any data recorded.

E.g. A quick check was done on a PH1000 with the current output connected to an ammeter. The results recorded were 8.48 mA in pH 4.00 buffer and 14.82 mA in a pH 10.01 buffer.

Expected readings are calculated by:

Expected reading (mA) = (pH value x sensor output range / sensor pH range) + 4

i.e. Expected reading in pH 4 buffer =  $(4.00 \times (20-4) / (14-0)) + 4$   
= 8.57 mA

Expected reading in pH10 buffer =  $(10.01 \times (20-4) / (14-0)) + 4$   
= 15.44 mA

The gain correction can be calculated with the formula-

$$\text{Gain correction} = \frac{(\text{expected reading at high pH} - \text{expected reading at low pH})}{(\text{Reading at high pH} - \text{Reading at low pH})}$$

$$\begin{aligned}\text{Gain correction} &= (15.44 - 8.57) / (14.82 - 8.48) \\ &= 1.0836\end{aligned}$$

The offset correction can be calculated with the formula –

$$\text{Offset correction} = \text{expected value at low pH} - (\text{reading at low pH} \times \text{Gain correction})$$

$$\begin{aligned}\text{Offset correction} &= 8.57 - (8.48 \times 1.0836) \\ &= -0.619\end{aligned}$$

### 3.3 Installation

#### Shipping and Storage Cap



The PH1000 sensor is shipped with a rubber cap covering the pH electrode. This cap contains a small amount of pH 4 buffer and is in place to prevent the electrode from drying out during shipping and storage. **This cap must be removed prior to installation.** The cap should be retained so that it can be placed back on the electrode if the sensor is to be removed.

#### Field Deployment Considerations

The sensor head should always be completely submerged and positioned such that the possibility of air bubbles becoming entrapped on the electrode is minimised. Bubbles may cause errors if trapped. The sensor should be periodically inspected for fouling, and can be cleaned with fresh water and damp cloth. In marine environments crustaceans may need removal at regular intervals.

The body should always be fully immersed under the water to ensure the electronic module is at water temperature and to avoid any possible anodic/cathodic action taking place on the stainless body due to the oxygen difference across the boundary. Care should be taken if clamps are to be attached to the Stainless Steel body as the depletion of oxygen to the clamp/probe interface can cause corrosion due to anodic/cathodic action. For applications where stainless steel might be corroded it is recommended that the Acetal body be specified.

Sensors should generally be installed such that they can be easily and safely removed for cleaning and servicing. For environmental applications the sensor can often be mounted inside a section of PVC or steel pipe which enters the water body. The sensor can then be slid down inside the pipe until the sensor head just protrudes into the water body. This provides a high degree of protection for the sensor from environmental (sunlight, heat, flood debris etc) as well as from other influences such as Cattle,

vandalism etc. Most sediment transport occurs during storm events and flood conditions. Protection from floating debris damage is an important consideration along with adequate tethering of sensors.

Another widespread application of the sensors is to hang the sensors in the water body from a fixed structure or a floating buoy or pontoon. Generally in lakes or estuary applications the sensor can hang on its own cable, and easily retrieved for routine servicing.

### ***Cabling Considerations***

Care should be taken with installation and field servicing to ensure the cable is not subjected to persistent pulling snagging or severe compression. Cyclic loading of the cable should also be avoided through careful sensor deployment. Additional stilling wells or mounting brackets may be required to prevent sensor movement which may cause long term cable movement. Where cable runs are required which may be subject to environmental effects (heat, water movement, sunlight, flood debris etc) it is advisable to protect the sensor cable inside a slightly larger diameter conduit such as PVC, steel or polyethylene. This also allows the sensor cable to be pulled out – should a sensor change-over be required at the site. Maximum cable runs up to several hundred meters are possible without affecting electrical signals. The maximum cable length is dependant on the capability of the com port of the computer. Most computers should be capable of driving a 150 to 200m cable length.

### ***Typical Sensor Installations***

1. Edge of river/stream/lake embankment.
2. Side of boat/vessel.
3. Mounted within a stilling well off stream from main flow.
4. Mounted within drainage channels/pipes.
5. Suspended from dam walls or floating pontoon.
6. Sensor anchored to bed of lake/stream.

### ***Field Installation must ensure:***

- The sensor is anchored or held in position or located so it is not subject to any movement during normal operations.
- Sensor is protected from direct sunlight to avoid high temperature fluctuations
- Sensor is protected against high turbulence and possible debris loading during flow events
- The rubber storage cap is removed from the electrode

### ***Other Considerations***

Environmental compatibility should be checked before using the sensors and advice sought from Greenspan if any doubt exists. The sensor utilises some 316 stainless components that are suitable in a majority of situations but care should be taken against possible corrosion in high Chloride, Sulphate or Ferric solutions. The body should always be totally immersed under the water to ensure that the sensor is at water temperature and to also avoid any possible anodic/cathodic action taking place on the components at the water-air interface. If using clamps to mount the sensor – these should be of a type that evenly clamps the sensor body without excessive loading that could damage the sensor body.

## ***Guidelines for cleaning equipment***



The sensor and electrode may be cleaned using a soft cloth, mild detergents and warm water. If the sensor shows signs of marine growth a light biocide can be used to clean and kill any biological growth on the sensor. Sensors are supplied with a removable shroud that provides added protection for the electrode. This shroud unscrews for cleaning.

The sensor must be cleaned with the electrode in place. The sensor will be damaged by the ingress of water or other solutions into the electrode cavity. The electrode should only be removed when it is to be replaced. To assist in cleaning the glass electrode, the PH1000 is provided with a special tool. A few drops of a mild detergent can be placed in the cleaning tool which is then pushed over the glass pH bulb. Carefully rotating the tool by hand should remove most contamination. For heavy scaling or contamination a cleaner that contains some abrasive can be used. Rinse the electrode in clean water prior to installing sensor or checking calibration.

## ***Replacing Reference Protection Ring***



The reference protection ring can be removed and/or replaced (with the electrode still in place in the sensor) by unscrewing the plastic shroud on the end of the electrode. The white protection ring can be slid off the glass electrode.

## ***Electrode storage***



PH electrodes should be stored in such a way to prevent the electrode from drying out. A rubber cap is provided with the electrode for this purpose. A small amount of liquid should be placed in this cap prior to the cap being placed over the electrode. PH 4 buffer is the preferred storage solution however, tap water should be used if not available. Soaking in pH 4 buffer solution overnight may restore some electrodes that have been allowed to dry out, although some calibration drift can be expected until the electrode is fully re-hydrated.

## **Electrode replacement**

PH electrodes will require periodic replacement. How long they last is dependant on the environment and to a lesser extent the cleaning and maintenance. In a typical environment, electrode life between 1 and 3 years is quite possible. The electrode is sealed in its mount with 2 o rings on the electrode body. The electrode is removed by unscrewing it from the housing. Spanner flats are machined on the electrode to assist.



Prior to removal of an electrode the sensor must be thoroughly cleaned and dried. Care must be exercised to ensure that no liquids or other contaminants enter the electrode cavity. The pH sensor should not be stored or transported without an electrode. Immersion of a PH1000 without an electrode will likely cause damage to the sensor.

With the old electrode removed the electrode cavity should be carefully inspected. A small ring of contamination may be present at the mouth of the cavity. If present, this contamination should be removed with a clean cloth in such a way so that the contamination is not pushed inside the cavity. Do not flush the cavity with water or other cleaning solvents as damage to the electronics inside the sensor could occur.

Once clean, a new electrode can be screwed into the electrode mount. It is recommended that the calibration of the new electrode be checked prior to re-deployment.



## 4 Appendix A -Additional Information

### 4.1 Specifications



Measurement technique	Gel-filled glass electrode with internal Ag/AgCl reference Field Replaceable electrode
Sensor range (factory calibrated)	0-14 pH (4mA = 0pH, 20mA = 14pH)
Sensor output	Analogue 4-20mA
Accuracy	+/-0.2pH (+/- 0.2mA)
Cable type	Polyurethane sheathed cable, OD 8mm, with 3mm vent tube, moulded entry, bare wire connection
Cable lengths	10, 20, 30, 50, 100, 150 m (32, 65, 100, 165, 325, 490 ft)
Power supply	10-14 VDC (at sensor)
Reverse polarity protection	Yes
Surge current protected	To 2kV
Warm up time to stable reading	2 sec
Current consumption	10mA to 30mA while turned on
Operating temperature	0-50°C
Storage temperature	-5°C - +60°C
Depth rating (water column)	100m
Weight	500g plus cable weight (665g per 10m length)
Dimensions (L x OD)	384.20mm x 47mm (15.13" x 1.85")
Wetted materials	UPVC, acetal, 316 passivated stainless steel, polyurethane, viton

## 4.2 Quick Start Guide – PH1000 Sensor

Procedures for connecting and configuring the Greenspan PH1000 Sensor. Further details in the Sensor Manual included on the Greenspan CD.

### You should have received:

- Certificate of Conformance Sheet
- CD Manuals & Software
- Greenspan PH1000 Sensor
- Cable Options per order



*PH1000 provides pH measurement as a calibrated analogue outputs 4-20mA, over the Sensor Full Scale range. Sensor requires a nominal 12V DC power supply, but will operate from 10-14V DC*

## 4.3 Certificate of Conformance



PH1000 pH Sensor

# CERTIFICATE of CONFORMANCE

**Customer:** "Click here & type Customer name"  
**Model No.** PH1000 pH Sensor (Tyco Environmental Systems material # 700-4400)  
**Sales Order Reference:** "Click here & type SO Reference"  
**Serial Number:** "Click here & type S/N"

### Product Information



PH1000 pH Sensor

<b>Range</b>	pH	0-14 pH
<b>Accuracy</b>		See note 7
<b>Cable Length</b>		"Click here & type Cable length" M
<b>Ext Supply Voltage</b>		11 - 14 VDC
<b>Power</b>	<b>+ve</b>	Red
	<b>Gnd</b>	Green
	<b>Output</b>	Blue
	<b>Shield</b>	Yellow/Green
<b>Output</b>		4- 20mA
<b>Connection Code</b>		BW3

### User Notes

1. The sensor is protected against reverse polarity.
2. Do not attempt to dismantle the sensor as it will void the warranty. Contact your agent for technical advice.
3. A spare Reference Protection Ring has been provided with your sensor. This a porous Teflon impregnated with a KCl gel.
4. The sensor utilises a Gel filled pH probe. They will generally last for 12-18 months in the field. A field replaceable pH probe is available to service your sensor at the deployment site. Contact your sales representative for details.
5. The glass bulb located inside the black end shroud should be kept clean. Marine growth or silt deposit may affect sensor accuracy. Cleaning schedules are determined by site conditions and initially weekly visits are suggested for visual inspections. The pH probe should be cleaned with a weak soap solution using provided cleaning tool. For severe build up use a gentle abrasive cleaner such as Ajax
6. The pH sensor bulb should always be kept moist. Ensure the cap provided has a small quantity of pH4 buffer inside when storing or transporting.
7. The sensor turn on time is two seconds and is factory set.
8. Sensor Accuracy - **pH:**  $\pm 0.2\text{pH}$

Inspected By: \_\_\_\_\_ 11 March 2010

Manufactured By:  
 Tyco Environmental Systems (Greenspan Analytical Plant)  
 22 Palmerin Street  
 WARWICK QLD 4370  
 AUSTRALIA Phone: + 61 (0)7 4660188



## 4.4 Engineering Note – Detachable Cables

### ENGINEERING NOTE

RELEASE DATE: 28/4/2008

SUBJECT: Connecting and Disconnecting the Detachable Cable for Greenspan Sensors (packaged in 47mm & 65mm Tubes)

#### IMPORTANT NOTE:

You have been supplied with:

2 x O-Rings (Greenspan part number 011-OR16X2.5)


2 x Grub Screws (Greenspan part number 512-M4X6SS316P)

1 x Allen Key


**Prior to connecting your sensors to the detachable cable, please ensure that the detachable cable entry has been fitted with O-Rings supplied as detailed above. Ensure O-Ring grooves are greased and then slide O-Rings and fit into the recessed grooves.**




Refer to the following instructions for **connecting** your sensor to the detachable cable



1. The sensor connector is exposed and ready to connect to the cable connector.



2. Align the cable connector with the sensor connector and push together. The connector is polarized and will only fit together one way.



3. Once connectors are completely pushed together, the rocker arm on the sensor connector will clip in with the cable connector and secure the connection.



4. Carefully feed the connection back into the sensor tube taking caution not to crimp or damage any wiring. Ensure the O-Rings have been fitted to the moulded cable entry as directed at the beginning of this document



5. Twist carefully to align the grub screw holes. Firmly push the tube back onto the moulded cable entry



7. Screw in two (2) 512-M4X6SS316P grub screws to secure the cable and the sensor with supplied Allen key



8. The cable is now securely fitted to the sensor.


Refer to the following instructions for **disconnecting** your sensor to the detachable cable




1. Sensor with cable attached and ready for cable disconnection.




2. Remove the two (2) grub screws (Greenspan part #512-M4X6SS316P) using the Allen key supplied. (N.B Allen key is supplied in spare parts kit with original packaging)



3. CAREFULLY pull tube away from cable mould. Once the tube has been pulled away from cable mould, the connector will be revealed as pictured left.



4. Release locking device on the connector by gently pinching rocker arm on the cable connector and pull apart.



5. The sensor and cable are now detached.

**Note:** When the cable is not connected to a sensor, please ensure the supplied vent cap is fitted to the vent tube (if applicable)